



Volumetric flasks, trapezoidal

BLAUBRAND®, class A, DE-M marking

Boro 3.3. DIN EN ISO 1042. Calibrated to contain (TC, In). Incl. one batch certificate. Pack of 2.

Capacity ml	Error limit ± ml	Neck size	with PP stopper Cat. No.	with glass stopper Cat. No.
1	0.025	7/16	364 01	364 12
2	0.025	7/16	364 02	364 13
5	0.025	7/16	364 03	364 14
10 W	0.04	10/19	364 04	364 15
20	0.04	10/19	364 06	364 17
25	0.04	10/19	364 07	364 18
50	0.06	12/21	364 08	364 19

W = Wide neck

Volumetric flasks

BLAUBRAND®, class A, DE-M marking

Boro 3.3. DIN EN ISO 1042. Calibrated to contain (TC, In). Incl. one batch certificate. 5 to 500 ml: Pack of 2; 1000 to 10000 ml: Pack of 1.



Capacity ml	Error limit ± ml	Neck size	with PP stopper Cat. No.	with glass stopper Cat. No.
5 W	0.04	10/19	372 38	372 56
10 W	0.04	10/19	372 43	372 67
20	0.04	10/19	372 46	372 57
20 W	0.06	12/21	372 45	372 68
25	0.04	10/19	372 47	372 58
25 W	0.06	12/21	372 93	372 94
50	0.06	12/21	372 48	372 59
50 W	0.10	14/23	372 90	372 88
100	0.10	12/21	372 49	372 60
100	0.10	14/23	372 91	372 89
200	0.15	14/23	372 50	372 61
250	0.15	14/23	372 51	372 62
500	0.25	19/26	372 52	372 63
1000	0.4	24/29	372 53	372 64
1000 W	0.6	29/32	372 34	—
2000	0.6	29/32	372 54	372 65
5000	1.2	34/35	372 55	372 66
10000*	2.0	45/40	372 36	—

W = Wide neck * in addition to the ISO range

Connect With Us

